PINAL • COUNTY
wide open opportunity

Thermometer Calibration Log

(Calibration is the process of resetting a thermometer to read accurately at a known temperature)

Calibration Instructions:



- ☐ Fill a large glass with ice and water then stir.
- ☐ Immerse thermometer stem 2 inches in water until needle stops moving. Hold the stem away from the bottom and sides to avoid error.
- ☐ Rotate the nut *(under head)* until the pointer reads **32°F** while in ice water by using a mini wrench or thermometer cover adjusting hole.
- ☐ Check Calibration weekly and when thermometer is dropped.
- ☐ Sanitize the thermometer before using by wiping with an alcohol swab or by washing the stem in soap, rinsing, and sanitizing.

Date	Temperature Before Calibration	Temperature After Calibration	Calibrated by	Comments

ENVIRONMENTAL HEALTH SERVICES

971 Jason Lopez Cir. Bldg D Florence, AZ 85132 **T** 520-866-6864 **FREE** 866-960-0633 **F** 520-866-7066 www.pinalcountyaz.gov